



Gallios FlowCheck 2015-10-19
Flow-Check Pro - 10 color - Acquisition

[Ungated] FS




Gate	Number	X-Mean	HP	X-CV
All	10,432	524.03		1.77
FCblue@2.0	5,000	840.78		0.60
FCred	2,515	306.69		1.11
FCviolet	2,293	122.02		1.77

[FCblue@2.0] FL1




Gate	Number	X-Mean	HP	X-CV
All	5,000	552.78		0.77
FL1@2.0	4,967	555.59		0.77

[FCblue@2.0] FL2




Gate	Number	X-Mean	HP	X-CV
All	5,000	527.33		0.72
FL2@2.0	4,964	529.94		0.72

[FCblue@2.0] FL3




Gate	Number	X-Mean	HP	X-CV
All	5,000	531.38		0.72
FL3@2.0	4,965	534.05		0.72


Gallios FlowCheck 2015-10-19
Flow-Check Pro - 10 color - Acquisition




Gate	Number	X-AMean	HP	X-CV
All	5,000	493.92		0.77
FL4@2.0	4,965	496.32		0.77



Gate	Number	X-AMean	HP	X-CV
All	5,000	551.40		1.16
FL5@2.5	4,960	554.11		1.16




Gate	Number	X-AMean	HP	X-CV
All	2,515	502.19		0.92
FL6@3.0	2,483	508.48		0.92




Gate	Number	X-AMean	HP	X-CV
All	2,515	455.83		1.20
FL7@3.0	2,482	461.56		1.20


Gallios FlowCheck 2015-10-19
Flow-Check Pro - 10 color - Acquisition



Gate	Number	X-Mean	HP	X-CV
All	2,515	509.62		1.24
FL8@3.0	2,477	516.06		1.24



Gate	Number	X-Mean	HP	X-CV
All	2,293	469.99		1.22
FL9@4.0	2,199	480.00		1.22



Gate	Number	X-Mean	HP	X-CV
All	2,293	505.23		1.31
FL10@4.0	2,252	511.66		1.31