







Gate	Number	X-AMean	HP	X-CV
All	11,371	483.75		2.18
FCblue@2.0	5,000	809.75		0.89
FCred	2,937	299.76		1.28
FCviolet	2,643	118.68		2.18



Gate	Number	X-AMean	HP	X-CV
All	5,000	529.89		0.80
FL1@2.0	4,933	532.38		0.80




Gate	Number	X-AMean	HP	X-CV
All	5,000	520.63		0.98
FL2@2.0	4,849	522.53		0.98



Gate	Number	X-AMean	HP	X-CV
All	5,000	516.21		0.99
FL3@2.0	4,878	518.20		0.99


[FCblue@2.0] FL4



FL4

Gate	Number	X-Mean	HP	X-CV
All	5,000	466.80		1.18
FL4@2.0	4,953	469.02		1.18


[FCblue@2.0] FL5



FL5

Gate	Number	X-Mean	HP	X-CV
All	5,000	529.34		1.12
FL5@2.5	4,754	530.51		1.12


[FCred] FL6



FL6

Gate	Number	X-Mean	HP	X-CV
All	2,937	487.11		1.39
FL6@3.0	2,915	490.04		1.39


[FCred] FL7



FL7


Gate	Number	X-Mean	HP	X-CV
All	2,937	441.72		1.90
FL7@3.0	2,887	444.67		1.90

[FCred] FL8




Gate	Number	X-AMean	HP	X-CV
All	2,937	493.20		1.54
FL8@3.0	2,915	496.09		1.54

[FCviolet] FL9



Gate	Number	X-AMean	HP	X-CV
All	2,643	498.26		1.50
FL9@4.0	2,615	502.17		1.50

[FCviolet] FL10



Gate	Number	X-AMean	HP	X-CV
All	2,643	508.45		1.42
FL10@4.0	2,611	511.63		1.42