



[Ungated] FS




Gate	Number	X-Mean	HP	X-CV
All	10,492	582.14		1.90
FCblue@2.0	5,000	940.38		0.63
FCred	2,445	340.68		1.25
FCviolet	2,418	135.72		1.90

[FCblue@2.0] FL1




Gate	Number	X-Mean	HP	X-CV
All	5,000	507.29		0.84
FL1@2.0	4,980	508.64		0.84

[FCblue@2.0] FL2




Gate	Number	X-Mean	HP	X-CV
All	5,000	502.44		1.01
FL2@2.0	4,981	503.78		1.01

[FCblue@2.0] FL3




Gate	Number	X-Mean	HP	X-CV
All	5,000	517.71		0.89
FL3@2.0	4,975	519.07		0.89

[FCblue@2.0] FL4




Gate	Number	X-Mean	HP	X-CV
All	5,000	484.72	1.05	
FL4@2.0	4,964	485.64	1.05	

[FCblue@2.0] FL5




Gate	Number	X-Mean	HP	X-CV
All	5,000	510.12	1.00	
FL5@2.5	4,978	511.27	1.00	

[FCred] FL6




Gate	Number	X-Mean	HP	X-CV
All	2,445	495.68	0.77	
FL6@3.0	2,430	498.38	0.77	

[FCred] FL7




Gate	Number	X-Mean	HP	X-CV
All	2,445	511.03	1.08	
FL7@3.0	2,429	513.79	1.08	

[FCred] FL8




Gate	Number	X-Mean	HP	X-CV
All	2,445	512.56	1.00	
FL8@3.0	2,424	515.35	1.00	

[FCviolet] FL9



Gate	Number	X-Mean	HP	X-CV
All	2,418	507.74	1.66	
FL9@4.0	2,402	510.77	1.66	

[FCviolet] FL10



Gate	Number	X-Mean	HP	X-CV
All	2,418	507.61	1.48	
FL10@4.0	2,383	510.93	1.48	