




Gallios FlowCheck 2016-05-17
Flow-Check Pro - 10 color - Acquisition




Gate	Number	X-Mean	HP	X-CV
All	5,396	467.37		1.81
FCblue@2.0	2,500	791.03		0.54
FCred	1,319	253.29		1.01
FCviolet	1,317	93.51		1.81




Gate	Number	X-Mean	HP	X-CV
All	2,500	510.25		1.74
FL1@2.0	2,495	510.42		1.74




Gate	Number	X-Mean	HP	X-CV
All	2,500	496.29		2.19
FL2@2.0	2,493	496.39		2.19




Gate	Number	X-Mean	HP	X-CV
All	2,500	477.89		1.76
FL3@2.0	2,497	478.13		1.76




Gate	Number	X-Mean	HP	X-CV
All	2,500	509.06	1.90	
FL4@2.0	2,488	508.68	1.90	




Gate	Number	X-Mean	HP	X-CV
All	2,500	519.79	2.22	
FL5@2.5	2,498	519.62	2.22	




Gate	Number	X-Mean	HP	X-CV
All	1,319	509.58	1.00	
FL6@3.0	1,310	511.76	1.00	




Gate	Number	X-Mean	HP	X-CV
All	1,319	504.72	1.01	
FL7@3.0	1,311	506.75	1.01	



Gate	Number	X-Mean	HP	X-CV
All	1,319	498.63		1.19
FL8@3.0	1,310	500.77		1.19



Gate	Number	X-Mean	HP	X-CV
All	1,317	479.91		0.94
FL9@4.0	1,277	493.61		0.94



Gate	Number	X-Mean	HP	X-CV
All	1,317	476.61		0.70
FL10@4.0	1,283	487.99		0.70