




Gallios FlowCheck 2016-06-13
Flow-Check Pro - 10 color - Acquisition




Gate	Number	X-AMean	HP	X-CV
All	5,386	470.09	1.81	
FCblue@2.0	2,500	789.44	0.70	
FCred	1,339	253.78	1.17	
FCviolet	1,293	93.38	1.81	




Gate	Number	X-AMean	HP	X-CV
All	2,500	512.77	1.57	
FL1@2.0	2,496	512.96	1.57	




Gate	Number	X-AMean	HP	X-CV
All	2,500	495.51	2.33	
FL2@2.0	2,496	495.69	2.33	




Gate	Number	X-AMean	HP	X-CV
All	2,500	508.82	1.83	
FL3@2.0	2,498	509.05	1.83	




Gate	Number	X-Mean	HP	X-CV
All	2,500	513.10		1.89
FL4@2.0	2,488	512.71		1.89




Gate	Number	X-Mean	HP	X-CV
All	2,500	519.79		2.12
FL5@2.5	2,498	519.62		2.12




Gate	Number	X-Mean	HP	X-CV
All	1,339	513.03		0.99
FL6@3.0	1,330	514.47		0.99




Gate	Number	X-Mean	HP	X-CV
All	1,339	508.32		1.17
FL7@3.0	1,325	509.44		1.17



Gate	Number	X-Mean	HP	X-CV
All	1,339	503.11		1.10
FL8@3.0	1,331	504.50		1.10



Gate	Number	X-Mean	HP	X-CV
All	1,293	482.86		1.37
FL9@4.0	1,256	493.17		1.37



Gate	Number	X-Mean	HP	X-CV
All	1,293	483.02		0.69
FL10@4.0	1,258	493.01		0.69