



Gallios Flow Check 2016-06-27
Flow-Check Pro - 10 color - Acquisition

[Ungated] FS




Gate	Number	X-Mean	HP	X-CV
All	5,219	480.99		2.23
FCblue@2.0	2,500	792.21		0.65
FCred	1,298	255.23		1.17
FCviolet	1,177	93.88		2.23

[FCblue@2.0] FL1




Gate	Number	X-Mean	HP	X-CV
All	2,500	514.63		1.74
FL1@2.0	2,490	514.70		1.74

[FCblue@2.0] FL2



Gate	Number	X-Mean	HP	X-CV
All	2,500	498.68		2.57
FL2@2.0	2,493	498.82		2.57


[FCblue@2.0] FL3



Gate	Number	X-Mean	HP	X-CV
All	2,500	510.17		2.01
FL3@2.0	2,496	510.24		2.01


Gallios Flow Check 2016-06-27
Flow-Check Pro - 10 color - Acquisition

[FCblue@2.0] FL4




Gate	Number	X-AMean	HP	X-CV
All	2,500	509.89		1.93
FL4@2.0	2,488	509.91		1.93

[FCblue@2.0] FL5




Gate	Number	X-AMean	HP	X-CV
All	2,500	522.90		2.20
FL5@2.5	2,498	522.96		2.20

[FCred] FL6




Gate	Number	X-AMean	HP	X-CV
All	1,298	516.04		0.91
FL6@3.0	1,291	516.77		0.91

[FCred] FL7




Gate	Number	X-AMean	HP	X-CV
All	1,298	510.51		1.00
FL7@3.0	1,285	510.63		1.00


Gallios Flow Check 2016-06-27
Flow-Check Pro - 10 color - Acquisition



Gate	Number	X-Mean	HP	X-CV
All	1,298	505.73		1.10
FL8@3.0	1,291	506.47		1.10



Gate	Number	X-Mean	HP	X-CV
All	1,177	487.13		1.11
FL9@4.0	1,148	496.01		1.11



Gate	Number	X-Mean	HP	X-CV
All	1,177	484.31		0.95
FL10@4.0	1,148	492.93		0.95