



Gallios FlowCheck
Flow-Check Pro - 10 color - Acquisition

[Ungated] FS




[FCblue@2.0] FL1




Gate	Number	X-Mean	HP	X-CV
All	5,283	549.31		1.86
FCblue@2.0	2,500	905.11		0.80
FCred	1,365	313.40		0.95
FCviolet	1,196	112.84		1.86

Gate	Number	X-Mean	HP	X-CV
All	2,500	511.39		1.06
FL1@2.0	2,496	511.29		1.06

[FCblue@2.0] FL2



[FCblue@2.0] FL3




Gate	Number	X-Mean	HP	X-CV
All	2,500	514.76		1.14
FL2@2.0	2,469	514.13		1.14

Gate	Number	X-Mean	HP	X-CV
All	2,500	510.64		1.16
FL3@2.0	2,496	510.51		1.16


Gallios FlowCheck
Flow-Check Pro - 10 color - Acquisition

[FCblue@2.0] FL4




Gate	Number	X-Mean	HP	X-CV
All	2,500	505.18		1.08
FL4@2.0	2,496	505.06		1.08

[FCblue@2.0] FL5




Gate	Number	X-Mean	HP	X-CV
All	2,500	514.42		1.32
FL5@2.5	2,500	514.42		1.32

[FCred] FL6



Gate	Number	X-Mean	HP	X-CV
All	1,365	514.44		0.74
FL6@3.0	1,363	515.19		0.74


[FCred] FL7



Gate	Number	X-Mean	HP	X-CV
All	1,365	509.86		0.92
FL7@3.0	1,360	510.52		0.92


Gallios FlowCheck
Flow-Check Pro - 10 color - Acquisition

[FCred] FL8




Gate	Number	X-AMean	HP	X-CV
All	1,365	502.84		1.19
FL8@3.0	1,363	503.57		1.19

[FCviolet] FL9



Gate	Number	X-AMean	HP	X-CV
All	1,196	507.64		1.56
FL9@4.0	1,185	512.35		1.56

[FCviolet] FL10



Gate	Number	X-AMean	HP	X-CV
All	1,196	495.13		1.27
FL10@4.0	1,185	499.68		1.27