



Gallios FlowCheck 2016-10-17
Flow-Check Pro - 10 color - Acquisition

[Ungated] FS




Gate	Number	X-Mean	HP	X-CV
All	4,394	608.07		1.90
FCblue@2.0	2,500	887.42		1.01
FCred	952	306.41		1.24
FCviolet	764	110.62		1.90

[FCblue@2.0] FL1




Gate	Number	X-Mean	HP	X-CV
All	2,500	500.16		0.93
FL1@2.0	2,498	500.02		0.93

[FCblue@2.0] FL2




Gate	Number	X-Mean	HP	X-CV
All	2,500	495.45		0.94
FL2@2.0	2,497	495.24		0.94

[FCblue@2.0] FL3



Gate	Number	X-Mean	HP	X-CV
All	2,500	492.21		0.78
FL3@2.0	2,498	492.07		0.78


[FCblue@2.0] FL4



FL4

Gate	Number	X-AMean	HP	X-CV
All	2,500	493.26		0.95
FL4@2.0	2,498	493.11		0.95


[FCblue@2.0] FL5



FL5

Gate	Number	X-AMean	HP	X-CV
All	2,500	502.91		1.02
FL5@2.5	2,498	502.76		1.02


[FCred] FL6



FL6

Gate	Number	X-AMean	HP	X-CV
All	952	515.35		0.58
FL6@3.0	951	515.90		0.58


[FCred] FL7




FL7

Gate	Number	X-AMean	HP	X-CV
All	952	503.94		0.76
FL7@3.0	949	504.38		0.76


Gallios FlowCheck 2016-10-17
Flow-Check Pro - 10 color - Acquisition



Gate	Number	X-AMean	HP	X-CV
All	952	487.04		1.40
FL8@3.0	949	487.64		1.40



Gate	Number	X-AMean	HP	X-CV
All	764	516.36		1.27
FL9@4.0	736	530.52		1.27



Gate	Number	X-AMean	HP	X-CV
All	764	504.94		1.36
FL10@4.0	735	505.36		1.36