




Gallios FlowCheck 2016-11-28
Flow-Check Pro - 10 color - Acquisition




Gate	Number	X-Mean	HP	X-CV
All	4,821	586.56		1.83
FCblue@2.0	2,500	916.26		0.98
FCred	1,144	316.22		0.94
FCviolet	975	114.43		1.83



Gate	Number	X-Mean	HP	X-CV
All	2,500	505.25		1.24
FL1@2.0	2,490	505.12		1.24




Gate	Number	X-Mean	HP	X-CV
All	2,500	506.92		1.24
FL2@2.0	2,477	506.32		1.24



Gate	Number	X-Mean	HP	X-CV
All	2,500	503.89		1.59
FL3@2.0	2,496	503.73		1.59


[FCblue@2.0] FL4



FL4

Gate	Number	X-AMean	HP	X-CV
All	2,500	499.30		1.35
FL4@2.0	2,487	499.13		1.35


[FCblue@2.0] FL5



FL5

Gate	Number	X-AMean	HP	X-CV
All	2,500	507.52		1.17
FL5@2.5	2,498	507.43		1.17


[FCred] FL6



FL6


Gate	Number	X-AMean	HP	X-CV
All	1,144	490.87		0.86
FL6@3.0	1,136	492.78		0.86

[FCred] FL7




FL7


Gate	Number	X-AMean	HP	X-CV
All	1,144	486.18		1.05
FL7@3.0	1,140	487.89		1.05



Gate	Number	X-Mean	HP	X-CV
All	1,144	481.72		1.06
FL8@3.0	1,135	483.64		1.06



Gate	Number	X-Mean	HP	X-CV
All	975	505.61		0.49
FL9@4.0	952	515.50		0.49



Gate	Number	X-Mean	HP	X-CV
All	975	491.62		0.77
FL10@4.0	952	496.23		0.77