



[Ungated] FS




Gate	Number	X-Mean	HP	X-CV
All	4,835	589.48		2.22
FCblue@2.0	2,500	920.49		1.06
FCred	1,057	316.44		1.48
FCviolet	1,040	115.12		2.22

[FCblue@2.0] FL1




Gate	Number	X-Mean	HP	X-CV
All	2,500	517.41		1.13
FL1@2.0	2,487	517.42		1.13

[FCblue@2.0] FL2



Gate	Number	X-Mean	HP	X-CV
All	2,500	516.92		1.14
FL2@2.0	2,460	516.37		1.14


[FCblue@2.0] FL3



Gate	Number	X-Mean	HP	X-CV
All	2,500	514.31		1.23
FL3@2.0	2,497	514.52		1.23


Gallios FlowCheck 2017-03-06
Flow-Check Pro - 10 color - Acquisition

[FCblue@2.0] FL4




Gate	Number	X-AMean	HP	X-CV
All	2,500	502.84		1.26
FL4@2.0	2,491	502.85		1.26

[FCblue@2.0] FL5




Gate	Number	X-AMean	HP	X-CV
All	2,500	511.06		1.74
FL5@2.5	2,497	511.09		1.74

[FCred] FL6




Gate	Number	X-AMean	HP	X-CV
All	1,057	497.17		1.01
FL6@3.0	1,042	503.90		1.01

[FCred] FL7



Gate	Number	X-AMean	HP	X-CV
All	1,057	488.59		1.29
FL7@3.0	1,043	495.15		1.29


[FCred] FL8



FL8

Gate	Number	X-Mean	HP	X-CV
All	1,057	482.57		1.23
FL8@3.0	1,037	489.28		1.23


[FCviolet] FL9



FL9

Gate	Number	X-Mean	HP	X-CV
All	1,040	517.05		0.86
FL9@4.0	1,002	532.85		0.86

[FCviolet] FL10



FL10

Gate	Number	X-Mean	HP	X-CV
All	1,040	504.53		1.24
FL10@4.0	999	512.28		1.24