



Gallios FlowCheck
Flow-Check Pro - 10 color - Acquisition

[Ungated] FS




Gate	Number	X-AMean	HP	X-CV
All	4,728	605.68		1.44
FCblue@2.0	2,500	929.62		0.96
FCred	1,025	320.22		1.20
FCviolet	960	116.46		1.44

[FCblue@2.0] FL1




Gate	Number	X-AMean	HP	X-CV
All	2,500	514.31		1.07
FL1@2.0	2,491	514.37		1.07

[FCblue@2.0] FL2



Gate	Number	X-AMean	HP	X-CV
All	2,500	517.10		1.13
FL2@2.0	2,462	516.55		1.13


[FCblue@2.0] FL3



Gate	Number	X-AMean	HP	X-CV
All	2,500	514.87		1.15
FL3@2.0	2,494	514.93		1.15


Gallios FlowCheck
Flow-Check Pro - 10 color - Acquisition

[FCblue@2.0] FL4




Gate	Number	X-Mean	HP	X-CV
All	2,500	502.24		1.10
FL4@2.0	2,488	501.74		1.10

[FCblue@2.0] FL5




Gate	Number	X-Mean	HP	X-CV
All	2,500	515.24		1.40
FL5@2.5	2,498	514.83		1.40

[FCred] FL6



Gate	Number	X-Mean	HP	X-CV
All	1,025	520.34		0.81
FL6@3.0	1,011	523.78		0.81


[FCred] FL7



Gate	Number	X-Mean	HP	X-CV
All	1,025	508.24		1.09
FL7@3.0	1,012	511.49		1.09


Gallios FlowCheck
Flow-Check Pro - 10 color - Acquisition

[FCred] FL8




Gate	Number	X-AMean	HP	X-CV
All	1,025	503.41	1.01	
FL8@3.0	1,017	506.96	1.01	

[FCviolet] FL9



Gate	Number	X-AMean	HP	X-CV
All	960	528.51	0.98	
FL9@4.0	913	551.95	0.98	

[FCviolet] FL10



Gate	Number	X-AMean	HP	X-CV
All	960	516.26	1.22	
FL10@4.0	910	525.42	1.22	