




Gallios FlowCheck 2017-11-13
Flow-Check Pro - 10 color - Report Sheet 1




Gate	HP X-CV
FCblue@2.0	1.02
FCred	0.94
FCviolet	1.81




Gate	HP X-CV
FL1@2.0	0.89




Gate	HP X-CV
FL2@2.0	0.98




Gate	HP X-CV
FL3@2.0	1.05




Gate	HP X-CV
FL4@2.0	1.03




Gate	HP X-CV
FL5@2.5	1.26




Gate	HP X-CV
FL6@3.0	0.98




Gate	HP X-CV
FL7@3.0	1.02



Gate	HP X-CV
FL8@3.0	0.91



Gate	HP X-CV
FL9@4.0	1.17



Gate	HP X-CV
FL10@4.0	1.14

Date of data set acquisition... 13-NOV-2017