




Gallios FlowCheck 2017-12-18
Flow-Check Pro - 10 color - Report Sheet 1




Gate	HP X-CV
FCblue@2.0	0.95
FCred	0.92
FCviolet	1.43




Gate	HP X-CV
FL1@2.0	0.97




Gate	HP X-CV
FL2@2.0	1.13




Gate	HP X-CV
FL3@2.0	0.88




Gate	HP X-CV
FL4@2.0	1.04




Gate	HP X-CV
FL5@2.5	1.26




Gate	HP X-CV
FL6@3.0	0.67




Gate	HP X-CV
FL7@3.0	1.10



Gate	HP X-CV
FL8@3.0	0.83



Gate	HP X-CV
FL9@4.0	0.91



Gate	HP X-CV
FL10@4.0	1.13

Date of data set acquisition... 18-DEC-2017