





Gallios FlowCheck 2016-07-25
Flow-Check Pro - 10 color - Acquisition



Gate	Number	X-AMean	HP	X-CV
All	5,554	459.96		2.26
FCblue@2.0	2,500	790.90		0.76
FCred	1,327	254.50		1.01
FCviolet	1,421	94.08		2.26


Gate	Number	X-AMean	HP	X-CV
All	2,500	517.35		1.71
FL1@2.0	2,492	517.38		1.71




Gate	Number	X-AMean	HP	X-CV
All	2,500	495.96		2.04
FL2@2.0	2,494	496.07		2.04

Gate	Number	X-AMean	HP	X-CV
All	2,500	514.59		1.71
FL3@2.0	2,496	514.61		1.71


Gallios FlowCheck 2016-07-25
 Flow-Check Pro - 10 color - Acquisition




Gate	Number	X-AMean	HP	X-CV
All	2,500	517.13		1.86
FL4@2.0	2,480	516.93		1.86



Gate	Number	X-AMean	HP	X-CV
All	2,500	525.77		1.94
FL5@2.5	2,498	525.84		1.94




Gate	Number	X-AMean	HP	X-CV
All	1,327	521.38		0.81
FL6@3.0	1,319	523.79		0.81




Gate	Number	X-AMean	HP	X-CV
All	1,327	514.79		1.15
FL7@3.0	1,298	516.55		1.15

[FCred] FL8




Gate	Number	X-AMean	HP	X-CV
All	1,327	508.37		0.83
FL8@3.0	1,318	510.67		0.83

[FCviolet] FL9



Gate	Number	X-AMean	HP	X-CV
All	1,421	471.27		1.12
FL9@4.0	1,349	493.08		1.12

[FCviolet] FL10



Gate	Number	X-AMean	HP	X-CV
All	1,421	482.31		0.77
FL10@4.0	1,373	496.68		0.77